O ViEh¤ SiÂSnpVAXQuto中間र्यवाStErilization PE: Animal Care Personnel

SCOPE:

Facility Manager and Technical Staff **RESPONSIBILITY:**

- 4. Items to be sterilized should be **placed into the autoclave** so that steam can uniformly flow between items and so that no air pockets are formed between or around the items.
- 5. **Small items** to be autoclaved **should be bagged or wrapped** in heat-resistant plastic, paper, or cloth.
- 6. **Sterilized supplies must be labeled** with date autoclaved/processed. Date of expiration is determined using the following criteria:
 - a. Sterile packs/items should be stored in clean, dry enclosed cabinets and drawers free from moisture and dust.
 - b. Sterile packs prepared by research or technical staff, when stored in enclosed cabinets/drawers in a dust-free, insect-free, and well-ventilated area, can be considered sterile for 1 year providing the packaging remains dry and intact.
 - c. Sterile packs prepared by research or technical staff stored on open shelves have a reduced safe shelf life of 1 month providing the packaging remains dry and intact.
 - d. The manufacturer generally stamps commercially prepared sterilized packs with an expiration date.
 - e. A shelf life of **3 year** is suggested for commercially prepared packs without an expiration date, providing the pack is stored in an enclosed cabinet/drawer, remains dry, and packaging is intact. **Label with date of receipt.**
 - f. The quality of the packaging material, storage and transport conditions, and the amount of handling all contribute to maintaining sterility of the package and its contents.
 - g. Any package that is wet, torn, dropped on the floor, damaged in any way, or pastital date of exempit kitkeage that (tsi)/0.7((ce) that,) ITytiurs wet, torn, dr(by)3.7(r)0.7(es)-1.7(ear6)

| SOP #1006.6 | |
|----------------|-----------|
| Autoclave Ster | ilization |
| Effective 4/22 | |
| Page 4 of 4 | |
| - | |

| Approved: Date: | |
|-----------------|--|
|-----------------|--|